



ASSAM GOVERNMENT-IITG HEALTHCARE FOUNDATION

CIN: U85300AS2022NPL023684 REG

Pre-Bid Meeting Summary

Date: 16.07.2025

RFB reference no: AGIHF/RFB/SUPPLY/2025-26/003

Brief Description of Procurement: Selection of eligible firms/ manufacturers for supply of components for a single low-field MRI unit to Assam Government – IIT Guwahati Healthcare Foundation.

Date and Time for Pre-Bid meeting: July 11, 2025 11 AM

Venue of Pre-Bid Meeting: Conference Room of AGIHF, 5th Floor, Research Block, IIT Guwahati.

The following Bidder attended the pre-bid meeting:

SL No	Name of prospective bidder/ firm	Name
1	Scimertech	Dr Manoj Nimbalkar
2	Medirays Imaging Private Limited	Mr Rajesh Nair
3	Medikabazar United Imaging Healthcare	Mr Partha Pratim Deka

The panel as appointed by the Chief Operating Officer, AGIHF were present in the pre-bid meeting to provide clarification to the queries. In order to bring the uniformity and clarity on various terms and conditions as mentioned in the tender document, the queries received from the above-mentioned bidder during pre-bid meeting were discussed and clarified. The clarifications arrived thereof are indicated hereunder:

Medikabazar United Imaging Healthcare				
Sl. No.	NIT Description	Document reference/ Page No.	Clarifications sought	Remarks by AGIHF
1.	Technical Specification	Page 8 of 14	Can the specification be changed to 1.5 tesla/ 1500 mT from the existing 0.05 tesla/ 50 mT	No, we require the field strength to be 0.05 tesla/ 50 mT
Medirays Imaging Private Limited				
2.	Delivery Time	Page 5 of 14	What is the delivery time	Delivery to be done within 60 days of issue of Purchase Order. Please

Regd. Ofc: 3rd Floor, Centre for Nanotechnology, IIT Guwahati, North Guwahati, Kamrup Rural, Guwahati-781039, Assam, India

Corp. Ofc: 5th Floor, Research Park, IIT Guwahati, North Guwahati, Kamrup Rural, Guwahati-781039, Assam, India

Email: info@agihf.org Ph: 0361 2910509



ASSAM GOVERNMENT-IITG HEALTHCARE FOUNDATION

CIN: U85300AS2022NPL023684 REG

				refer corrigendum dated: 16.07.2025
3.	General		The product to be delivered will be classified as medical device or an industrial device	Since the product is intended for research purpose only, the products are to be classified and delivered as an industrial device.
4.	Certification	Page 3 of 14	What are the certifications and required?	1. ISO 9001:2015 – Quality Management, 2. NABL-accredited test report confirming electrical safety of the device for user protection, as per relevant industrial standards
5.	General		Is import allowed.	The tender does not restrict any imports.